

GSO ISO 13485:2016

ISO 13485:2016

مواصفة قياسية خليجية

الأجهزة الطبية – نظم إدارة الجودة – المتطلبات التنظيمية

Gulf Standard

Medical devices -- Quality management systems --
Requirements for regulatory purposes

ICS: 03.100 ; 11.040.01

اللغة الإنجليزية – English Language

الطبعة 1 – 22/12/2016 – Edition 1



© GCC Standardization Organization , www.gso.org.sa

جميع الحقوق محفوظة لهيئة التقييس لدول مجلس التعاون لدول الخليج العربية ©



هذه الوثيقة محمية بالقانون
COPYRIGHT PROTECTED DOCUMENT

جميع الحقوق محفوظة لهيئة لتقييس لدول مجلس التعاون لدول الخليج العربية. ما لم ينص على خلاف ذلك، لا يجوز إعادة إنتاج أي جزء من هذا المنشور أو استخدامه بأي شكل آخر أو بأي وسيلة، إلكترونية أو ميكانيكية، بما في ذلك التصوير، أو النشر على الإنترنت أو الإنترنت، دون إذن كتابي مسبق من هيئة التقييس لدول مجلس التعاون لدول الخليج العربية على العنوان أدناه أو الدول الأعضاء في الهيئة داخل بلدانهم.

All rights reserved to GCC Standardization Organization (GSO). Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either GSO at the address below or GSO's member body in the country of the requester.

Outreach Dept, Knowledge Management Section

GCC Standardization Organization (GSO)

P.O.Box: 85245 Riyadh 12511

Kingdom of Saudi Arabia

إدارة التواصل، قسم إدارة المعرفة

هيئة التقييس لدول مجلس التعاون لدول الخليج العربية (GSO)

ص.ب: 85245 الرياض 12511

المملكة العربية السعودية

Tel:	+966115208000	هاتف:
Fax:	+966115208010	فاكس:
E-Mail:	csc@gso.org.sa	بريد إلكتروني:
Website:	www.gso.org.sa	موقع ويب:

تقديم خليجي

هيئة التقييس لدول مجلس التعاون لدول الخليج العربية هيئة إقليمية تضم في عضويتها أجهزة التقييس الوطنية في الدول الأعضاء ، ومن مهام الهيئة إعداد المواصفات القياسية واللوائح الفنية الخليجية بواسطة لجان فنية متخصصة.

قرر المجلس الفني لهيئة التقييس لدول مجلس التعاون لدول الخليج العربية في الاجتماع رقم (40) الذي عقد بتاريخ 1439/10/03 هـ ، الموافق 2016/12/22 م اعتماد المواصفة القياسية الخليجية رقم GSO ISO 13485:2016 وعنوانها "الأجهزة الطبية – نظم إدارة الجودة – المتطلبات التنظيمية" باللغة الانجليزية التي تم دراستها وتبنيها عن المنظمة الدولية للتقييس رقم ISO 13485:2016 بدون إدخال تعديلات فنية عليها ضمن برنامج عمل اللجنة الفنية الخليجية لمواصفات الأجهزة والمستلزمات الطبية (TC11) المدرجة في خطة المملكة العربية السعودية .

Gulf Foreword

GCC Standardization Organization (GSO) is a regional organization which consists of the national standardization bodies of GCC member states. One of GSO main functions is to issue gulf standards and technical regulations through specialized technical committees.

GSO Technical Council in its meeting number (40) held on 03/10/1439 H , which corresponds to 22/12/2016 has approved gulf standard No GSO ISO 13485:2016 titled "Medical devices -- Quality management systems -- Requirements for regulatory purposes " in English language which was studied and adopted from International Organization for Standardization Standard ISO 13485:2016 without any technical modifications through the technical program of GSO Technical Committee for Medical Devices and Supplies (TC11) in SAUDI ARABIA schedule .

INTERNATIONAL STANDARD

ISO
13485

Third edition
2016-03-01

Medical devices — Quality management systems — Requirements for regulatory purposes

*Dispositifs médicaux — Systèmes de management de la qualité —
Exigences à des fins réglementaires*



Reference number
ISO 13485:2016 (E)

© ISO 2016 © GSO 2016



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016 — All rights reserved, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.ch
Web www.iso.ch

Contents	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Quality management system	
4.1 General requirements	
4.2 Documentation requirements	
4.2.1 General	
4.2.2 Quality manual	
4.2.3 Medical device file	
4.2.4 Control of documents	
4.2.5 Control of records	
5 Management responsibility	
5.1 Management commitment	
5.2 Customer focus	
5.3 Quality policy	
5.4 Planning	
5.4.1 Quality objectives	
5.4.2 Quality management system planning	
5.5 Responsibility, authority and communication	
5.5.1 Responsibility and authority	
5.5.2 Management representative	
5.5.3 Internal communication	
5.6 Management review	
5.6.1 General	
5.6.2 Review input	
5.6.3 Review output	
6 Resource management	
6.1 Provision of resources	
6.2 Human resources	
6.3 Infrastructure	
6.4 Work environment and contamination control	
6.4.1 Work environment	
6.4.2 Contamination control	
7 Product realization	
7.1 Planning of product realization	
7.2 Customer-related processes	
7.2.1 Determination of requirements related to product	
7.2.2 Review of requirements related to product	
7.2.3 Communication	
7.3 Design and development	
7.3.1 General	
7.3.2 Design and development planning	
7.3.3 Design and development inputs	
7.3.4 Design and development outputs	
7.3.5 Design and development review	
7.3.6 Design and development verification	
7.3.7 Design and development validation	

	7.3.8	Design and development transfer	
	7.3.9	Control of design and development changes	
	7.3.10	Design and development files	
7.4		Purchasing	
	7.4.1	Purchasing process	
	7.4.2	Purchasing information	
	7.4.3	Verification of purchased product	
7.5		Production and service provision	
	7.5.1	Control of production and service provision	
	7.5.2	Cleanliness of product	
	7.5.3	Installation activities	
	7.5.4	Servicing activities	
	7.5.5	Particular requirements for sterile medical devices	
	7.5.6	Validation of processes for production and service provision	
	7.5.7	Particular requirements for validation of processes for sterilization and sterile barrier systems	
	7.5.8	Identification	
	7.5.9	Traceability	
	7.5.9.1	General	
	7.5.9.2	Particular requirements for implantable medical devices	
	7.5.10	Customer property	
	7.5.11	Preservation of product	
7.6		Control of monitoring and measuring equipment	
8		Measurement, analysis and improvement	
	8.1	General	
	8.2	Monitoring and measurement	
	8.2.1	Feedback	
	8.2.2	Complaint handling	
	8.2.3	Reporting to regulatory authorities	
	8.2.4	Internal audit	
	8.2.5	Monitoring and measurement of processes	
	8.2.6	Monitoring and measurement of product	
8.3		Control of nonconforming product	
	8.3.1	General	
	8.3.2	Actions in response to nonconforming product detected before delivery	
	8.3.3	Actions in response to nonconforming product detected after delivery	
	8.3.4	Rework	
8.4		Analysis of data	
8.5		Improvement	
	8.5.1	General	
	8.5.2	Corrective action	
	8.5.3	Preventive action	
		Annex A (informative) Comparison of content between ISO 13485:2003 and ISO 13485:2016	
		Annex B (informative) Correspondence between ISO 13485:2016 and ISO 9001:2015	
		Bibliography	7

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is Technical Committee ISO/TC 210, *Quality management and corresponding general aspects for medical devices*.

This third edition of ISO 13485 cancels and replaces the second edition (ISO 13485:2003) and ISO/TR 14969:2004, which have been technically revised. It also incorporates the Technical Corrigendum ISO 13485:2003/Cor.1:2009. A summary of the changes incorporated into this edition compared with the previous edition is given in Annex A.

Introduction

0.1 General

This International Standard specifies requirements for a quality management system that can be used by an organization involved in one or more stages of the life-cycle of a medical device, including design and development, production, storage and distribution, installation, servicing and final decommissioning and disposal of medical devices, and design and development, or provision of associated activities (e.g. technical support). The requirements in this International Standard can also be used by suppliers or other external parties providing product (e.g. raw materials, components, subassemblies, medical devices, sterilization services, calibration services, distribution services, maintenance services) to such organizations. The supplier or external party can voluntarily choose to conform to the requirements of this International Standard or can be required by contract to conform.

Several jurisdictions have regulatory requirements for the application of quality management systems by organizations with a variety of roles in the supply chain for medical devices. Consequently, this International Standard expects that the organization:

- identifies its role(s) under applicable regulatory requirements;
- identifies the regulatory requirements that apply to its activities under these roles;
- incorporates these applicable regulatory requirements within its quality management system.

The definitions in applicable regulatory requirements differ from nation to nation and region to region. The organization needs to understand how the definitions in this International Standard will be interpreted in light of regulatory definitions in the jurisdictions in which the medical devices are made available.

This International Standard can also be used by internal and external parties, including certification bodies, to assess the organization's ability to meet customer and regulatory requirements applicable to the quality management system and the organization's own requirements. It is emphasized that the quality management system requirements specified in this International Standard are complementary to the technical requirements for product that are necessary to meet customer and applicable regulatory requirements for safety and performance.

The adoption of a quality management system is a strategic decision of an organization. The design and implementation of an organization's quality management system is influenced by the:

- a) organizational environment, changes in that environment, and the influence that the organizational environment has on the conformity of the medical devices;
- b) organization's varying needs;
- c) organization's particular objectives;
- d) product the organization provides;
- e) processes the organization employs;
- f) organization's size and organizational structure;
- g) regulatory requirements applicable to the organization's activities.

It is not the intent of this International Standard to imply the need for uniformity in the structure of different quality management systems, uniformity of documentation or alignment of documentation to the clause structure of this International Standard.

There is a wide variety of medical devices and some of the particular requirements of this International Standard only apply to named groups of medical devices. These groups are defined in [Clause 3](#).

0.2 Clarification of concepts

In this International Standard, the following terms or phrases are used in the context described below.

- When a requirement is qualified by the phrase “as appropriate”, it is deemed to be appropriate unless the organization can justify otherwise. A requirement is considered appropriate if it is necessary for:
 - product to meet requirements;
 - compliance with applicable regulatory requirements;
 - the organization to carry out corrective action;
 - the organization to manage risks.
- When the term “risk” is used, the application of the term within the scope of this International Standard pertains to safety or performance requirements of the medical device or meeting applicable regulatory requirements.
- When a requirement is required to be “documented”, it is also required to be established, implemented and maintained.
- When the term “product” is used, it can also mean “service”. Product applies to output that is intended for, or required by, a customer, or any intended output resulting from a product realization process.
- When the term “regulatory requirements” is used, it encompasses requirements contained in any law applicable to the user of this International Standard (e.g. statutes, regulations, ordinances or directives). The application of the term “regulatory requirements” is limited to requirements for the quality management system and the safety or performance of the medical device.

In this International Standard, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

Information marked as “NOTE” is for guidance in understanding or clarifying the associated requirement.

0.3 Process approach

This International Standard is based on a process approach to quality management. Any activity that receives input and converts it to output can be considered as a process. Often the output from one process directly forms the input to the next process.

For an organization to function effectively, it needs to identify and manage numerous linked processes. The application of a system of processes within an organization, together with the identification and interactions of these processes, and their management to produce the desired outcome, can be referred to as the “process approach.”

When used within a quality management system, such an approach emphasizes the importance of:

- a) understanding and meeting requirements;
- b) considering processes in terms of added value;

- c) obtaining results of process performance and effectiveness;
- d) improving processes based on objective measurement.

0.4 Relationship with ISO 9001

While this is a stand-alone standard, it is based on ISO 9001:2008, which has been superseded by ISO 9001:2015. For the convenience of users, Annex B shows the correspondence between this International Standard and ISO 9001:2015.

This International Standard is intended to facilitate global alignment of appropriate regulatory requirements for quality management systems applicable to organizations involved in one or more stages of the life-cycle of a medical device. This International Standard includes some particular requirements for organizations involved in the life-cycle of medical devices and excludes some of the requirements of ISO 9001 that are not appropriate as regulatory requirements. Because of these exclusions, organizations whose quality management systems conform to this International Standard cannot claim conformity to ISO 9001 unless their quality management system meets all the requirements of ISO 9001.

0.5 Compatibility with other management systems

This International Standard does not include requirements specific to other management systems, such as those particular to environmental management, occupational health and safety management, or financial management. However, this International Standard enables an organization to align or integrate its own quality management system with related management system requirements. It is possible for an organization to adapt its existing management system(s) in order to establish a quality management system that complies with the requirements of this International Standard.

Medical devices — Quality management systems — Requirements for regulatory purposes

1 Scope

This International Standard specifies requirements for a quality management system where an organization needs to demonstrate its ability to provide medical devices and related services that consistently meet customer and applicable regulatory requirements. Such organizations can be involved in one or more stages of the life-cycle, including design and development, production, storage and distribution, installation, or servicing of a medical device and design and development or provision of associated activities (e.g. technical support). This International Standard can also be used by suppliers or external parties that provide product, including quality management system-related services to such organizations.

Requirements of this International Standard are applicable to organizations regardless of their size and regardless of their type except where explicitly stated. Wherever requirements are specified as applying to medical devices, the requirements apply equally to associated services as supplied by the organization.

The processes required by this International Standard that are applicable to the organization, but are not performed by the organization, are the responsibility of the organization and are accounted for in the organization's quality management system by monitoring, maintaining, and controlling the processes.

If applicable regulatory requirements permit exclusions of design and development controls, this can be used as a justification for their exclusion from the quality management system. These regulatory requirements can provide alternative approaches that are to be addressed in the quality management system. It is the responsibility of the organization to ensure that claims of conformity to this International Standard reflect any exclusion of design and development controls.

If any requirement in Clauses 6, 7 or 8 of this International Standard is not applicable due to the activities undertaken by the organization or the nature of the medical device for which the quality management system is applied, the organization does not need to include such a requirement in its quality management system. For any clause that is determined to be not applicable, the organization records the justification as described in 4.2.2.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 9000:2015¹, *Quality management systems — Fundamentals and vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 9000:2015 and the following apply.

¹ Supersedes ISO 9000:2005.

3.1
advisory notice
notice issued by the organization, subsequent to delivery of the medical device, to provide supplementary information or to advise on action to be taken in the:

- use of a medical device,
- modification of a medical device,
- return of the medical device to the organization that supplied it, or
- destruction of a medical device

Note 1 to entry: Issuance of an advisory notice can be required to comply with applicable regulatory requirements.

3.2
authorized representative
natural or legal person established within a country or jurisdiction who has received a written mandate from the manufacturer to act on his behalf for specified tasks with regard to the latter's obligations under that country or jurisdiction's legislation

[SOURCE: GHTF/SG1/N055:2009, 5.2]

3.3
clinical evaluation
assessment and analysis of clinical data pertaining to a medical device to verify the clinical safety and performance of the device when used as intended by the manufacturer

[SOURCE: GHTF/SG5/N4:2010, Clause 4]

3.4
complaint
written, electronic or oral communication that alleges deficiencies related to the identity, quality, durability, reliability, usability, safety or performance of a medical device that has been released from the organization's control or related to a service that affects the performance of such medical devices

Note 1 to entry: This definition of "complaint" differs from the definition given in ISO 9000:2015.

3.5
distributor
natural or legal person in the supply chain who, on his own behalf, furthers the availability of a medical device to the end user

Note 1 to entry: More than one distributor may be involved in the supply chain.

Note 2 to entry: Persons in the supply chain involved in activities such as storage and transport on behalf of the manufacturer, importer or distributor, are not distributors under this definition.

[SOURCE: GHTF/SG1/N055:2009, 5.3]

3.6
implantable medical device
medical device which can only be removed by medical or surgical intervention and which is intended to:

- be totally or partially introduced into the human body or a natural orifice, or

- replace an epithelial surface or the surface of the eye, and
- remain after the procedure for at least 30 days

Note 1 to entry: This definition of implantable medical device includes active implantable medical device

3.7

importer

natural or legal person in the supply chain who is the first in a supply chain to make a medical device, manufactured in another country or jurisdiction, available in the country or jurisdiction where it is to be marketed

[SOURCE: GHTF/SG1/N055:2009, 5.4]

3.8

labelling

label, instructions for use, and any other information that is related to identification, technical description, intended purpose and proper use of the medical device, but excluding shipping documents

[SOURCE: GHTF/SG1/N70:2011, Clause 4]

3.9

life-cycle

all phases in the life of a medical device, from the initial conception to final decommissioning and disposal

[SOURCE: ISO 14971:2007, 2.7]

3.10

manufacturer

natural or legal person with responsibility for design and/or manufacture of a medical device with the intention of making the medical device available for use, under his name; whether or not such a medical device is designed and/or manufactured by that person himself or on his behalf by another person(s)

Note 1 to entry: This “natural or legal person” has ultimate legal responsibility for ensuring compliance with all applicable regulatory requirements for the medical devices in the countries or jurisdictions where it is intended to be made available or sold, unless this responsibility is specifically imposed on another person by the Regulatory Authority (RA) within that jurisdiction.

Note 2 to entry: The manufacturer’s responsibilities are described in other GHTF guidance documents. These responsibilities include meeting both pre-market requirements and post-market requirements, such as adverse event reporting and notification of corrective actions.

Note 3 to entry: “Design and/or manufacture”, as referred to in the above definition, may include specification development, production, fabrication, assembly, processing, packaging, repackaging, labelling, relabelling, sterilization, installation, or remanufacturing of a medical device; or putting a collection of devices, and possibly other products, together for a medical purpose.

Note 4 to entry: Any person who assembles or adapts a medical device that has already been supplied by another person for an individual patient, in accordance with the instructions for use, is not the manufacturer, provided the assembly or adaptation does not change the intended use of the medical device.

Note 5 to entry: Any person who changes the intended use of, or modifies, a medical device without acting on behalf of the original manufacturer and who makes it available for use under his own name, should be considered the manufacturer of the modified medical device.

Note 6 to entry: An authorized representative, distributor or importer who only adds its own address and contact details to the medical device or the packaging, without covering or changing the existing labelling, is not considered a manufacturer.

Note 7 to entry: To the extent that an accessory is subject to the regulatory requirements of a medical device, the person responsible for the design and/or manufacture of that accessory is considered to be a manufacturer.

[SOURCE: GHTF/SG1/N055:2009, 5.1]

3.11 medical device

instrument, apparatus, implement, machine, appliance, implant, reagent for *in vitro* use, software, material or other similar or related article, intended by the manufacturer to be used, alone or in combination, for human beings, for one or more of the specific medical purpose(s) of:

- diagnosis, prevention, monitoring, treatment or alleviation of disease;
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury;
- investigation, replacement, modification, or support of the anatomy or of a physiological process;
- supporting or sustaining life;
- control of conception;
- disinfection of medical devices;
- providing information by means of *in vitro* examination of specimens derived from the human body;

and does not achieve its primary intended action by pharmacological, immunological or metabolic means, in or on the human body, but which may be assisted in its intended function by such means

Note 1 to entry: Products which may be considered to be medical devices in some jurisdictions but not in others include: — disinfection substances; — aids for persons with disabilities; — devices incorporating animal and/or human tissues; — devices for *in vitro* fertilization or assisted reproduction technologies.

[SOURCE: GHTF/SG1/N071:2012, 5.1]

3.12 medical device family

group of medical devices manufactured by or for the same organization and having the same basic design and performance characteristics related to safety, intended use and function

3.13 performance evaluation

assessment and analysis of data to establish or verify the ability of an *in vitro* diagnostic medical device to achieve its intended use

3.14 post-market surveillance

systematic process to collect and analyse experience gained from medical devices that have been placed on the market

3.15**product**

result of a process

Note 1 to entry: There are four generic product categories, as follows: — services (e.g. transport); — software (e.g. computer program, dictionary); — hardware (e.g. engine mechanical part); — processed materials (e.g. lubricant). Many products comprise elements belonging to different generic product categories. Whether the product is then called service, software, hardware or processed material depends on the dominant element. For example, the offered product “automobile” consists of hardware (e.g. tyres), processed materials (e.g. fuel, cooling liquid), software (e.g. engine control software, driver's manual), and service (e.g. operating explanations given by the salesman).

Note 2 to entry: Service is the result of at least one activity necessarily performed at the interface between the supplier and customer and is generally intangible. Provision of a service can involve, for example, the following: — an activity performed on a customer-supplied tangible product (e.g. automobile to be repaired); — an activity performed on a customer-supplied intangible product (e.g. the income statement needed to prepare a tax return); — the delivery of an intangible product (e.g. the delivery of information in the context of knowledge transmission); — the creation of ambience for the customer (e.g. in hotels and restaurants). Software consists of information and is generally intangible and can be in the form of approaches, transactions or procedures. Hardware is generally tangible and its amount is a countable characteristic. Processed materials are generally tangible and their amount is a continuous characteristic. Hardware and processed materials often are referred to as goods.

Note 3 to entry: This definition of “product” differs from the definition given in ISO 9000:2015.

[SOURCE: ISO 9000:2005², 3.4.2, modified]

3.16**purchased product**

product provided by a party outside the organization's quality management system

Note 1 to entry: The provision of product does not necessarily infer a commercial or financial arrangement.

3.17**risk**

combination of the probability of occurrence of harm and the severity of that harm

Note 1 to entry: This definition of “risk” differs from the definition given in ISO 9000:2015.

[SOURCE: ISO 14971:2007, 2.16]

3.18**risk management**

systematic application of management policies, procedures and practices to the tasks of analysing, evaluating, controlling and monitoring risk

[SOURCE: ISO 14971:2007, 2.22]

3.19**sterile barrier system**

minimum package that prevents ingress of microorganisms and allows aseptic presentation of the product at the point of use

[SOURCE: ISO 11607-1:2006, 3.22]

² Superseded by ISO 9000:2015.

3.20

sterile medical device

medical device intended to meet the requirements for sterility

Note 1 to entry: The requirements for sterility of a medical device can be subject to applicable regulatory requirements or standards.

Bibliography

- [1] ISO 9001:2015³, *Quality management systems — Requirements*
- [2] ISO 10012, *Measurement management systems — Requirements for measurement processes and measuring equipment*
- [3] ISO 11607-1:2006, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*
- [4] ISO 11607-2, *Packaging for terminally sterilized medical devices — Part 2: Validation requirements for forming, sealing and assembly processes*
- [5] ISO 14644 (all parts), *Cleanrooms and associated controlled environments*
- [6] ISO 14698 (all parts), *Cleanrooms and associated controlled environments — Biocontamination control*
- [7] ISO 14971:2007, *Medical devices — Application of risk management to medical devices*
- [8] ISO 19011, *Guidelines for auditing management systems*
- [9] IEC 62366-1, *Medical devices — Part 1: Application of usability engineering to medical devices*
- [10] GHTF/SG1/N055:2009⁴, *Definition of the Terms “Manufacturer”, “Authorised Representative”, “Distributor” and “Importer”*
- [11] GHTF/SG5/N4:2010⁵, *Post-Market Clinical Follow-Up Studies*
- [12] GHTF/SG1/N70:2011⁶, *Label and Instructions for Use for Medical Devices*
- [13] GHTF/SG1/N071:2012⁷, *Definition of the Terms “Medical Device” and “In Vitro Diagnostic (IVD) Medical Device”*

³ Supersedes ISO 9001:2008.

⁴ Available from website: <http://www.imdrf.org/documents/doc-ghtf-sg1.asp>

⁵ Available from website: <http://www.imdrf.org/documents/doc-ghtf-sg5.asp>

⁶ Available from website: <http://www.imdrf.org/documents/doc-ghtf-sg1.asp>

⁷ Available from website: <http://www.imdrf.org/documents/doc-ghtf-sg1.asp>

